



**VSL**

**National  
Metrology  
Institute**


# Metrology for traceable protocols for elemental and oxidized mercury concentrations

Iris de Krom

CEM 2023 – 22<sup>nd</sup> September 2023 –  
Barcelona – Spain

Session: Mercury






MAKE  
MERCURY  
HISTORY




- **EN 14181** *Quality assurance of automated measuring systems*
  1. Requirements for the calibration and validation → **QAL 2**
    - With guidance specific for mercury measurements **EN 14884** *determination of total mercury – automated measuring systems*
      - 6.2.2 **Zero and span check** (EN 14181:2014, A.7): Elemental mercury shall be used for the independent span check provided that the reference material generator used by the test laboratory is calibrated with metrological traceability to the SI.
      - 6.2.3 **Linearity test** (EN 14181:2014, A.8 and Annex B): Linearity tests shall be performed by passing gaseous reference materials through the entire automated measuring systems
      - 6.2.4 **Response time** (EN 14181:2014, A.11): Response time tests shall be performed by passing gaseous reference materials through the entire automated measuring systems
      - 6.2.5 **Converter efficiency**: The converter efficiency shall be tested to confirm that oxidized mercury is converted to elemental mercury. Oxidized mercury reference materials shall be introduced (e.g.; HgCl<sub>2</sub>).
    - **EN 13211** *Manual method of determination of the concentration of total mercury SRM manual method*
      - based on wet chemistry
  2. Requirements for quality control → **QAL 3**
    - **Zero and span check**
- **US EPA** traceability protocol for qualification and certification of elemental and oxidised mercury gas generators

# Mercury gas generators



- Saturation gas generators (ISO 6145-9)
  - Manually with syringe injection (bell-jar)
  - Automatically
- Permeation gas generators (ISO 6145-10)
- Mercury amount fraction in cylinders (ISO 6142-1)
- Continuous injection (ISO 6145-4)
  - Based on vaporization of a mercury chloride solution




- Mercury concentration dependent upon vapour pressure equation used



# Traceability chain



# Oxidised mercury converter efficiency



$$\text{Converter efficiency} = \frac{R_C}{R_{\text{ref}}}$$

$R_C$	343
$R_{\text{ref}}$	392
$E_c$	88%

## Calibration protocol



- This protocol specifies the procedures for establishing **traceability to the SI** for the **quantitative output of mercury gas generators** that are employed in regulatory applications for emission or ambient air monitoring.
  
- This protocol provides methods for
  - Calibrating the output of a mercury gas generator by comparison with a reference standard;
  - Calculating the uncertainty of the mercury concentration generated with the gas generator in relation to the known uncertainty of the reference standard.






# Performance evaluation

- Goal
  - making sure the developed calibration protocol is fit for purpose
  - establish a benchmark
  - encouraging the use of the best available methods
- Characteristics
  - stabilisation period
  - short term drift
  - precision, i.e., reproducibility and repeatability of the concentration generated
  - linearity
  - bias




# Elemental mercury



- Stabilisation period: between 10 and 30 minutes
- Short term drift:  $\leq 2\%$
- Precision:  $\leq 2\%$
- Bias & linearity:



- HgCl<sub>2</sub> salt based mercury gas generator



- Liquid evaporative HgCl<sub>2</sub> gas generators



# HgCl<sub>2</sub> solution storage conditions



## Temperature:

Fridge

Ambient

Shock heat 60 °C



## Matrix:

0.1% v/v HCl

0.024% v/v HNO<sub>3</sub> + 0.0144% v/v HCl

0.125 µg/g HNO<sub>3</sub> + 0.125 µg/g HCl



## Bottle:

FEP

FLPE

Borosilicate



## Mercury concentration:

50 ng/g

200 ng/g

1000 ng/g

# HgCl<sub>2</sub> solution storage conditions



## Temperature:

Fridge

Ambient

Shock heat 60 °C



## Matrix:

0.1% v/v HCl

0.024% v/v HNO<sub>3</sub> + 0.0144% v/v HCl

0.125 µg/g HNO<sub>3</sub> + 0.125 µg/g HCl



## Bottle:

FEP

FLPE

Borosilicate



## Concentration:

50 ng/g

200 ng/g

1000 ng/g

## Conclusion

- SI-traceable calibration protocols for mercury gas generators used in the field
- Validation through performance evaluation of mercury gas generators on the market
- **Reports available online [www.SI-Hg.eu](http://www.SI-Hg.eu)**
- Calibration and measurement capabilities (CMCs) and ISO/IEC 17025 accreditation
  - Calibration of mercury analysers and mercury gas generators
  - Sampling of sorbent tubes

## Next steps

- NWIP in CEN/TC 264 “Air quality” WG8 “Measurement of total mercury emissions” (2024)
- Protocol converted into written documentary standard (2025)
- SI-traceable mercury measurement results for emission sources



# Acknowledgement



The EMPIR initiative is co-funded by the European Union's Horizon 2020 research and innovation programme and the EMPIR Participating States



National  
Metrology  
Institute



Thanks to our stakeholders and chief stakeholder: **David Graham Uniper**





**Contact Info:**  
**Iris de Krom**  
**[idekrom@VSL.nl](mailto:idekrom@VSL.nl)**  
**[www.SI-Hg.eu](http://www.SI-Hg.eu)**