



# 19NRM03 SI-Hg

## Validation results of the calibration of mercury gas generators

Workshop  
19 Sep 2023

**Timo Rajamäki** [timo.rajamaki@vtt.fi](mailto:timo.rajamaki@vtt.fi)  
**MIKES Metrology**  
**VTT Technical Research Centre of Finland Ltd**

# SI-Hg Objectives / Oxidised mercury



## Objective 2


*To validate a certification protocol for the certification of oxidised mercury ( $\text{Hg}^{\text{II}}$ ) gas generators used in the field for low mercury concentrations present in the atmosphere and higher concentrations from emission sources. The validation will include (1) metrological evaluation of state-of-the-art dual  $\text{Hg}^0$  and  $\text{Hg}^{\text{II}}$  analytical systems, (2) repeatability, reproducibility and uncertainty evaluation of the certification procedures at representative concentration levels extended to the low  $\text{ng}/\text{m}^3$  level.*

## Objective 3


*To organise a performance evaluation to gather data on the characteristics of at least three  $\text{Hg}^0$  and three  $\text{Hg}^{\text{II}}$  gas generators on the market.*

# Certification protocol for the certification of oxidized mercury gas generators



## CALIBRATION OF THE ANALYSER



## CALIBRATION OF THE GENERATOR




# Test setup including four HgII generators








# Traceability chain



# Dual analyzer calibration with traceably calibrated Hg<sup>0</sup> generator (secondary gas standard)






Hgtot correction factor	Hg0 calibration	Hg0 calibration + converter efficiency
Average	1,064	1,237
Standard deviation	0,011	0,013

# Evaporative generators, range 10µg/m<sup>3</sup>

Hg<sup>0</sup> < 1% of output

Data collected for:




- Accuracy
- Linearity
- Repeatability
- Reproducibility



# Evaporative generators, range $1\mu\text{g}/\text{m}^3$

$\text{Hg}^0 < 1\%$  of output  
Data collected for

- Accuracy
- Linearity
- Repeatability
- Reproducibility









# Evaporative generators, range $10\mu\text{g}/\text{m}^3$

$\text{Hg}^0 < 1\%$  of output  
Data collected for



- Accuracy
- Linearity
- Repeatability
- Reproducibility





# Evaporative generator with HgCl<sub>2</sub> and HgBr<sub>2</sub>, range 10µg/m<sup>3</sup>



# Salt based generator, range $10\mu\text{g}/\text{m}^3$



# Hg<sup>II</sup> generators, range 10µg/m<sup>3</sup>



## Validation results of the calibration of mercury gas generators

- Complicated setup with several parts and unknowns is needed for the process easily causing high uncertainties
- Process for traceable calibration of a  $\text{Hg}^{\text{II}}$  generator without direct measurement of  $\text{Hg}^{\text{II}}$  (or  $\text{Hg}^{\text{tot}}$ )
  - OK: Traceable calibration with  $\text{Hg}^0$
  - BUT: Two unknowns, generator and converter
- Fast response time of all system parts is highly beneficial in practical measurements
- The project outcome is several essential suggestions for the complete validation process